

Deutsche Akkreditierungsstelle GmbH

Entrusted according to Section 8 subsection 1 AkkStelleG in connection with Section 1 subsection 1 AkkStelleGBV

Signatory to the Multilateral Agreements of
EA, ILAC and IAF for Mutual Recognition

Accreditation



The Deutsche Akkreditierungsstelle GmbH attests that the calibration laboratory

Trigas FI GmbH
Erdinger Str. 2b, 85375 Neufahrn

is competent under the terms of DIN EN ISO/IEC 17025:2005 to carry out calibrations in the following fields:

Fluid quantities

- Liquid flow rate ^{a)}
- Volume of flowing liquids
- Mass of flowing liquids
- Gas flow rate
- Volume of flowing gases
- Mass of flowing gases

^{a)} also on-site calibration

The accreditation certificate shall only apply in connection with the notice of accreditation of 2018-03-05 with the accreditation number D-K-15149-01 and is valid until 2021-09-14. It comprises the cover sheet, the reverse side of the cover sheet and the following annex with a total of 4 pages.

Registration number of the certificate: **D-K -15149-01-00**



Deutsche Akkreditierungsstelle GmbH

Office Berlin
Spittelmarkt 10
10117 Berlin

Office Frankfurt am Main
Europa-Allee 52
60327 Frankfurt am Main

Office Braunschweig
Bundesallee 100
38116 Braunschweig

The publication of extracts of the accreditation certificate is subject to the prior written approval by Deutsche Akkreditierungsstelle GmbH (DAkKS). Exempted is the unchanged form of separate disseminations of the cover sheet by the conformity assessment body mentioned overleaf.

No impression shall be made that the accreditation also extends to fields beyond the scope of accreditation attested by DAkKS.

The accreditation was granted pursuant to the Act on the Accreditation Body (AkkStelleG) of 31 July 2009 (Federal Law Gazette I p. 2625) and the Regulation (EC) No 765/2008 of the European Parliament and of the Council of 9 July 2008 setting out the requirements for accreditation and market surveillance relating to the marketing of products (Official Journal of the European Union L 218 of 9 July 2008, p. 30). DAkKS is a signatory to the Multilateral Agreements for Mutual Recognition of the European co-operation for Accreditation (EA), International Accreditation Forum (IAF) and International Laboratory Accreditation Cooperation (ILAC). The signatories to these agreements recognise each other's accreditations.

The up-to-date state of membership can be retrieved from the following websites:

EA: www.european-accreditation.org

ILAC: www.ilac.org

IAF: www.iaf.nu

Deutsche Akkreditierungsstelle GmbH
German Accreditation Body

Annex to the Accreditation Certificate D-K-15149-01-00
according to DIN EN ISO/IEC 17025:2005

Period of validity: 2018-03-05 to 2021-09-14

Date of issue: 2018-03-05

Holder of certificate:

Trigas FI GmbH
Erdinger Str. 2b, 85375 Neufahrn

Head:

Athanasios Trigas

Deputy:

Harald Alexander

Cynthia Trigas

Lampros Michail

Accredited as calibration laboratory since: 2002-07-03

Calibrations in the fields:

Fluid quantities

- **Liquid flow rate** ^{a)}
- **Volume of flowing liquids**
- **Mass of flowing liquids**
- **Gas flow rate**
- **Volume of flowing gases**
- **Mass of flowing gases**

^{a)} also on-site calibration

Abbreviations used: see last page

Annex to the accreditation certificate D-K-15149-01-00

Permanent Laboratory

Measured quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability ¹⁾	Remarks
Liquid flow rate Volume flow rate dV/dt and volume V of flowing liquids	0.05 mL/min to 2000 L/min	Volumetric measurement (piston prover), Measured fluid: liquids with densities of 700 kg/m ³ to 1100 kg/m ³	0.04 %	Measuring instrument with frequency or analogue output or visual display
	10 L/min to 5000 L/min	Volumetric measurement (water flow calibrator) Measured fluid: water with density of 1000 kg/m ³	0.09 %	Measuring instrument with frequency output
			0.12 %	Measuring instrument with analogue output or visual display
Mass flow rate dm/dt and mass m of flowing liquids	0.04 g/min to 2000 kg/min	Volumetric measurement (piston prover), Unit conversion via density, Viscosities of 0,3 mm ² /s to 1600 mm ² /s	0.05 %	Measuring instrument with frequency or analogue output or visual display
	10 kg/min to 5000 kg/min	Volumetric measurement (water flow calibrator) Unit conversion via density of 1,0 mm ² /s	0.11 %	Measuring instrument with frequency output
			0.13 %	Measuring instrument with analogue output or visual display
Gas flow rate Volume flow rate dV/dt and volume V of flowing gases	Measuring range stated in standard conditions	Calibration object always downstream to flow standards		Measuring instrument with display of flow rate under actual condition or under standard condition Standard density in according to international accepted normative documents
	1 mL/min to 85 L/min	Laminar flow elements Calibration gas: dry air (dew point < -15 °C)	0.34 %	Measuring instrument with frequency or analogue output or visual display
	10 L/min to 20000 L/min	critical nozzle Calibration gas: dry air (dew point < -15 °C)	0.27 %	
	1 L/min to 1500 L/min	Bell prover Calibration gas: dry air (dew point < -15 °C)	0.26 %	
	20 mL/min to 4000 mL/min	Seal free piston prover, Calibration gas: dry air (dew point < -15 °C)	0.3 %	
	1 L/min to 700 L/min	Bell prover Calibration gases: - nitrogen N ₂ - argon Ar - helium He - and their mixtures	0.26 %	

¹⁾ The best measurement capabilities are stated according to DAkkS-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15149-01-00

Measured quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability ¹⁾	Remarks
Gas flow rate Volume flow rate dV/dt and volume V of flowing gases	Measuring range stated in standard conditions	Calibration object always downstream to flow standards		Measuring instrument with display of flow rate under actual condition or under standard condition Standard density in according to international accepted normative documents
	20 mL/min to 4000 mL/min	Seal free piston prover, Calibration gases: - nitrogen N_2 - argon Ar - helium He - and their mixtures	0.35 %	Measuring instrument with frequency or analogue output or visual display technically pure gases or mixtures with traced composition
	1 L/min to 300 L/min	Bell prover Calibration gases: - methane CH_4 - carbon dioxide CO_2 - propane C_3H_8 - and their mixtures - hydrogen H_2	0.26 %	
	20 mL/min to 4000 mL/min	Seal free piston prover, Calibration gases: - methane CH_4 - carbon dioxide CO_2 - propane C_3H_8 - and theirs mixtures - hydrogen H_2	0.35 %	
Mass flow rate dm/dt and mass m of flowing gases	1.3 mg/min to 110 g/min	Calibration object always downstream to flow standards Laminar flow elements Calibration gas: dry air (dew point < -15 °C)	0.36 %	
	12 g/min to 1440 kg/h	critical nozzle Calibration gas: dry air (dew point < -15 °C)	0.24 %	
	1290 mg/min to 1939 g/min	Bell prover Calibration gas: dry air (dew point < -15 °C)	0.27 %	
	25.8 mg/min to 5.17 g/min	Seal free piston prover, Calibration gas: dry air (dew point < -15 °C)	0.3 %	
Gas flow rate Mass flow rate dm/dt and mass m of flowing gases	1250 mg/min to 875 g/min 1784 mg/min to 1249 g/min 178 mg/min to 125 g/min	Bell prover Calibration gases: - nitrogen N_2 - argon Ar - helium He - and their mixtures	0.27 %	Measuring instrument with frequency or analogue output or visual display technically pure gases or mixtures with traced composition
	25 mg/min to 5.0 g/min 35 mg/min to 7.14 g/min 3.57 mg/min to 714 mg/min	Seal free piston prover, Calibration gases: - nitrogen N_2 - argon Ar - helium He - and their mixtures	0.35%	

¹⁾ The best measurement capabilities are stated according to DAkkS-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95 % and have a coverage factor of $k=2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15149-01-00

Measured quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability ¹⁾	Remarks
Gas flow rate Mass flow rate dm/dt and mass m of flowing gases	717 mg/min to 215 g/min 1970 mg/min to 593 g/min 2010 mg/min to 603 g/min 90 mg/min to 27 g/min	Bell prover Calibration gases: - methane CH ₄ - carbon dioxide CO ₂ - propane C ₃ H ₈ - and their mixtures - hydrogen H ₂	0.26 %	Measuring instrument with frequency or analogue output or visual display technically pure gases or mixtures with traced composition
	14.3 mg/min to 2.870 g/min 39.5 mg/min to 7.907 g/min 40 mg/min to 8.042 g/min 1.8 mg/min to 360 mg/min	seal free piston prover Calibration gases: - methane CH ₄ - carbon dioxide CO ₂ - propane C ₃ H ₈ - and their mixtures - hydrogen H ₂	0.35 %	

On-site calibration

Measured quantity / Calibration item	Range	Measurement conditions / procedure	Best measurement capability ¹⁾	Remarks
Liquid flow rate Volume flow rate dV/dt of flowing liquids	0.03 L/min to 2000 L/min	Volumetric measurement (Transfer standard) Reference turbines, coriolis, gear meter	0.09 %	DN 4 - DN 65
	10 L/min to 5000 L/min	Volumetric measurement (Transfer standard) Reference turbines, coriolis, gear meter	0.12 %	DN 20 - DN 150
Mass flow rate dm/dt of flowing liquids	0.025 kg/min to 2000 kg/min	Volumetric measurement (Transfer standard) Reference turbines, coriolis, gear meter Conversion by using density	0.09 %	DN 4 - DN 65
	10 kg/min to 5000 kg/min	Volumetric measurement (Transfer standard) Reference turbines, coriolis, gear meter Conversion by using density	0.12 %	DN 20 - DN 150

Abbreviations used:

DN Nominal diameter

¹⁾ The best measurement capabilities are stated according to DAkks-DKD-3 (EA-4/02). These are expanded uncertainties of measurement with a coverage probability of 95 % and have a coverage factor of $k=2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.